


BIOENDO

Xiamen Bioendo Technology Co., Ltd.



CERTIFICATE OF ANALYSIS

| Catalog Number: | G520030 | Lot Number: | 23081159 |
|--------------------------|--|--------------------------|--------------|
| Product: | Gel Clot Lyophilized Amebocyte Lysate Multi-test Vial (50 Tests/Vial) | Manufacture Date: | 08-Aug-2023 |
| Specifications | 5.2 ml/vial | Expiration Date: | 07-Aug -2025 |
| Test Items | Description | Results | |
| Appearance | White lyophilized powder | Pass | |
| Sensitivity | Within 50% to 200% of specified sensitivity | Pass | |
| Buffer capacity | pH 6.0-8.0 | Pass | |
| 4h negative control | Negative | Pass | |
| Matched Endotoxin | Lot No. 23045133, 90EU/vial | | |
| Storage | 2-8°C | | |

The sensitivity of reagent is 0.03 EU/ml when assay parallel with USP/China Pharmacopoeia Reference Standard Endotoxin. Lot No.:H0K354/150800-201601.

Reference:

United States Pharmacopoeia (USP), General Chapter <85>, Bacterial Endotoxins Test (BET) and China Pharmacopoeia <2020 version>, volume 4, general rules 1143.

Operator *Sandy xiao*

Date *2023.08.16*

QA Administrator *Jane Lin*

Date *2023.08.16*

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.bioendo.com>