



Endosafe[®]
Endotoxin Indicators (EVV100K)
Certificate of Quality

Each vial of Endotoxin Indicator Lot #: EVV01681 Expiration Date: Jun 2022

contains approximately 10,000 ng of lipopolysaccharide (derived from *E. coli* Strain 055:B5).

The biological activity is approximately 100,000 Endotoxin Units (EU) of lipopolysaccharide, as determined by serial two-fold dilution at the endpoint with Gel Clot U.S. Reference Lysate Lot # 15.

The absence of assay interfering substances was demonstrated by acceptable spike recovery in the range of 50% to 200%.

Qualified Analyst *Paul S. A.* Date 23 Jun 2020

Reviewed By *Lamar W. Saffert* Date 23 Jun 2020

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA