



CERTIFICATE OF ANALYSIS

Catalog Number:	ECV2500V	Lot Number:	26025123
Product:	Endotoxin Challenge Vial	Manufacture Date:	28-Feb-2026
Bacterial strain	<i>E.coli</i> O111: B4	Expiration Date:	27-Feb-2028

Contents	Description	Catalog No.	Lot No.
Endotoxin Challenge Vial	≥ 2500 EU/vial	ECV2500V	26025123
TEST (Method)	SPECIFICATIONS	Results	
Endotoxin Potency (EU/vial)	≥ 2500 EU/vial	Pass	
Storage	2-8°C		

Reconstitute ECV with 1.0 ml Water for BET, mix vigorously for at least 15 minutes on a vortex mixer.

Operator *Sandy Xiao*

Date *2026.03.04*

QA Administrator *Jane Lin*

Date *2026.07.04*