


BIOENDO

Xiamen Bioendo Technology Co., Ltd.



CERTIFICATE OF ANALYSIS

Catalog Number:	CSE10V	Lot Number:	24115185
Product:	Control Standard Endotoxin (CSE)	Manufacture Date:	07-Nov-2024
Bacterial strain	<i>E.coli</i> O111: B4	Expiration Date:	06-Nov-2027
Test Items	Description	Results	
Appearance	White lyophilized powder	Pass	
Potency (EU/vial)	90	Pass	
Storage	2-8°C		

Reconstitute CSE with 1.8ml Water for BET, mix vigorously for 5 minutes on a vortex mixer, obtaining 50 EU/ml standard stock solution. Reconstituted endotoxin solution of potency less than 20 EU/ml is stable for up to one day at 2-8°C, for endotoxin solution potency is equal or higher than 20 EU/ml, it is stable for at least one week at 2-8°C. The endotoxin level is 90 EU/vial when assay parallel with USP/China Pharmacopocia Reference Standard Endotoxin Lot No.: H0K354/150800-201601.

Reference:

United States Pharmacopocia (USP), General Chapter <85>, Bacterial Endotoxins Test (BET) and China Pharmacopocia <2020 version>, volume 4, general rules 1143.

Operator	Sandy xian	Date	2024.11.14
QA Administrator	Jane Lin	Date	2024.11.14

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.bioendo.com>