



## Endosafe<sup>®</sup> - PTS<sup>™</sup> Cartridges Certificate of Analysis



Reorder Code: PTS2001F Cartridge Lot #: 6601174 Expiration Date: Apr 2022



Calibration Code: 118752945338 RSE/CSE Ratio: 9 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 1-0.01 EU/mL Standard Curve Linearity: -0.998

Standard Curve Mean Reaction Times: 1.0 EU/mL 187 seconds

0.1 EU/mL 453 seconds

0.01 EU/mL 929 seconds

Archived Spike Concentration: 0.146 EU/mL Negative Control: Pass

**This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.**

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

**CAUTION: DO NOT FREEZE THE CARTRIDGES**

Qualified Analyst Bureau Date 11 Apr 2020

Reviewed By MEay Date 11 Apr 2020

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA