

BIOENDO



Xiamen Bioendo Technology Co., Ltd.

CERTIFICATE OF ANALYSIS

Catalog Number:	BH50	Lot Number:	22088143
Product:	50 mM Tris Buffer, pH7.0, 50ml/vial	Manufacture Date:	03-Aug-2022
		Expiration Date:	02-Aug-2024
Test (Method)	Specifications	Results	
Sterility	Negative	Negative	
Endotoxin Level (EU/ml)	< 0.005	Pass	
pH Test	7.0	7.0	
PPC Recovery (%)	50 ~ 150	97	
Storage	2-30°C		

Operator *Sandy Xiao*

Date *2022.08.10*

QA Administrator *Jane Wu*

Date *2022.08.10*

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.lalendotoxin.com>