

Endosafe[®]
LAL Grade Glassware (T100)
Certificate of Quality

Depyrogenated, 10 x 75 mm Tubes with Caps Lot Number P160518011 has been tested
by Quality Control and meets the following test specification:

Endotoxin Test Result: < 0.005 EU/device

Expiration Date: Jul 2022

Product Code: T100

Checked By *Pal GA* Date 31 Aug 2018

Reviewed By *J Yukin* Date 06 Sep 2018

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA