

CERTIFICATE OF ANALYSIS

Catalog Number: EC64405S **Lot Number:** 26050176
Product: Bioendo™ EC Endotoxin Test Kit
 (End-point Chromogenic Assay) **Expiration Date:** 04-May-2028
 64 Tests/Kit

Kit Contents	Description	Catalog No.	Lot No.
Lyophilized Amebocyte Lysate	1.7 ml/vial	EC17	26052015
Control Standard Endotoxin	20EU/vial Bacterial Strain: <i>E.coli</i> O111:B4	CSE10V	26055162
Chromogenic Substrate	1.7 ml/vial	ECCS17	26059025
Stop Reagent	20 ml/vial	ECSR20	26058159
Water for BET	50 ml/bottle	TRW50S	26066123

TEST (Method)	SPECIFICATIONS	Results
Linearity	$ r \geq 0.98$	Pass
Coefficient of Variation (%)	≤ 10	Pass
Endotoxin Range(EU/ml):0.1----1.0		
Enzyme Activity (OD)		
Blank (OD)	≤ 0.20	Pass
1EU/ml (OD)	≥ 0.75	Pass
Endotoxin Range(EU/ml):0.01----0.1		
Enzyme Activity (OD)		
Blank (OD)	≤ 0.30	Pass
0.1EU/ml (OD)	≥ 0.50	Pass

Recommended Reaction Times for the Assay Kit

Endotoxin Range (EU/ml)	T ₁ (minute)	T ₂ (minute)
0.01 --- 0.1	30	6
0.1 --- 1.0	15	6

Reconstitute CSE with 2.0 ml Water for BET. Mix vigorously for 5 minutes on a vortex mixer, obtaining 10 EU/ml standard stock solution.

The endotoxin level is 20EU/vial when assay parallel with USP Reference Standard Endotoxin Lot No.: R172R0

Operator *Ying Xiao*

Date *2016.06.03*

QA Administrator *Yanlin*

Date *2016.06.03*

Manufacturer Xiamen Bioendo Technology Co., Ltd.
 Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
 Website <http://www.lalendotoxin.com>