



Endosafe[®] - PTS[™] Cartridges Certificate of Analysis



Reorder Code: PTS2005F Cartridge Lot #: 6738134 Expiration Date: Aug 2022



Calibration Code: 514038377618 RSE/CSE Ratio: 11 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 5-0.05 EU/mL Standard Curve Linearity: -0.997

Standard Curve Mean Reaction Times: 5.0 EU/mL 140 seconds

0.5 EU/mL 369 seconds

0.05 EU/mL 783 seconds

Archived Spike Concentration: 0.96 EU/mL Negative Control: Pass

This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst Ben Nye

Date 27 Aug 2020

Reviewed By Shirley K. Keph

Date 27 Aug 2020

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA