

Certificated Of Analysis For The Product

The products described below were made in China.

GenFollower hereby certificates that the products listed below are made from virgin PP, produced by GenFollower Biotech under ISO9001: 2008 (No. 03612Q20667R0S) standard quality management system. All products are sterileized before shipment.

Cat. No.	Lot No.	Commodity descriptions and specifications	Manufacture date	Expire Date	QTY/CTN
TR5000-24	250628	5000 μ L / 5mL pipette tips, Nature, sterile, RNase and DNase free, 24 Tips/Rack, 6 Racks/Pack, 5 Packs/Case	2025/6/25	2030/6/25	24 x 6 x 5

They are manufactured of natural Polypropylene (PP) using diamond polished molds. Representative samples of the raw material have passed the USP 88 Biological Reactivity Tests, In Vivo for a Class VI plastic (Systemic Injection Test, Intracutaneous Test, Implantation Test) and USP 87 Biological Reactivity Tests, in Vitro for polymeric materials (Elution Test). In addition, Physico-chemical testing of this product met the USP limits. The products do not contain asbestos or Aluminum PCB (PCBA).

The sterilization process is EN 46002:1996 and EN 552:1994 (electron beam) certified to Medical Device Industry.

The pre-sterilized products are sterile as long as the protective plastic cover is unbroken or opened.

The expiry date for the products is 5 years from the manufacturing date.

The expiry date extends to the last day of the labelled expiry month.

Quality Assurance: 朱联斌

Certificate Date of Issue: 11th August 2025

绍兴锦帆生物科技有限公司
SHAOXING GENFOLLOWER BIOTECH CO., LTD.