



Endosafe[®] - PTS[™] Cartridges Certificate of Analysis



Reorder Code: PTS2001F Cartridge Lot #: 1639137 Expiration Date: May 2023



Calibration Code: 118454149998 RSE/CSE Ratio: 8 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 1-0.01 EU/mL Standard Curve Linearity: -0.998

Standard Curve Mean Reaction Times: 1.0 EU/mL 184 seconds

0.1 EU/mL 454 seconds

0.01 EU/mL 941 seconds

Archived Spike Concentration: 0.238 EU/mL Negative Control: Pass

This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst Jy Date 15 Jun 2021

Reviewed By C. Sun Date 15 Jun 2021

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA