



## Endosafe<sup>®</sup> - PTS<sup>™</sup> Cartridges Certificate of Analysis



Reorder Code: PTS2005F Cartridge Lot #: 8785129 Expiration Date: Oct 2020



Calibration Code: 514637996471 RSE/CSE Ratio: 12 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 5-0.05 EU/mL Standard Curve Linearity: -0.999

Standard Curve Mean Reaction Times: 5.0 EU/mL 146 seconds  
0.5 EU/mL 355 seconds  
0.05 EU/mL 779 seconds

Archived Spike Concentration: 0.84 EU/mL Negative Control: Pass

**This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.**

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

**CAUTION: DO NOT FREEZE THE CARTRIDGES**

Qualified Analyst Bm

Date 19 Oct 2018

Reviewed By [Signature]

Date 22 Oct 2018

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA