



# Endosafe<sup>®</sup> - PTS<sup>™</sup> Cartridges Certificate of Analysis



Reorder Code: PTS201F Cartridge Lot #: 1756142 Expiration Date: Sep 2023



Calibration Code: 613534581386 RSE/CSE Ratio: 10 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 10-0.1 EU/mL Standard Curve Linearity: -1.000

Standard Curve Mean Reaction Times: 10.0 EU/mL 135 seconds

1.0 EU/mL 332 seconds

0.1 EU/mL 745 seconds

Archived Spike Concentration: 0.66 EU/mL Negative Control: Pass

**This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.**

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

### CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst *[Signature]* Date 22 Sep 2021  
Reviewed By *[Signature]* Date 23 Sep 2021

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA