

Luoyang Fudau Biotech Co., Ltd

Certificate of Analysis

Inspection NO.: CJ- 202510001

Product name	Multi-wells cell culture plate	Specification	96wells, Non-TC treated
Item No.	C061096	Production batch	PE02251005
Package	Blister box, 1pcs/bag, 50bag/case	Validity Period	2028.09
Material	PS	Report date	2026.04.22
Material Grade	Meets the requirements of USP Class VI		
Test items	Standard	Test Result	
Exterior	Glossy surface, no deformation, no black spots, cracks, flakes, miscellaneous materials, oil stains, fingerprints, bubbles, flashes, etc.	Meet the requirements	
Structure size	Meet the requirements of the drawings	Meet the requirements	
Hydrophilia	After TC treatment, the contact angle is not greater than 45°	NA	
Package	Packing materials, quantity are correct, and packaging is intact	Meet the requirements	
Sterilize	radiation label should be red	Meet the requirements	
Bacterial endotoxin	Not more than 0.1EU/ml	Meet the requirements	
In vitro cytotoxin	Should meet the requirements of level 0	Meet the requirements	
Cell culture	Cells are in good condition and morphology	NA	
DNase detection	There should be no DNase	Meet the requirements	
RNase detection	There should be no RNase	Meet the requirements	
Testing base	QW/01-2004		
Conclusion	Qualified		

Inspector

张梦圆

Reviewer

张丽

Approver

白静