


BIOENDO



Xiamen Bioendo Technology Co., Ltd.

CERTIFICATE OF ANALYSIS

Catalog Number:	G520030	Lot Number:	22081149
Product:	Gel Clot Lyophilized Amebocyte Lysate Multi-test Vial (50 Tests/Vial)	Manufacture Date:	06-Aug-2022
Specifications	5.2 ml/vial	Expiration Date:	05-Aug-2024
Test Items	Description	Results	
Appearance	White lyophilized powder	Pass	
Sensitivity	Within 50% to 200% of specified sensitivity	Pass	
Buffer capacity	pH 6.0-8.0	Pass	
4h negative control	Negative	Pass	
Matched Endotoxin	Lot No. 22055125, 90EU/vial ; 22095149, 90EU/vial		
Storage	2-8°C		

The sensitivity of reagent is 0.03EU/ml when assay parallel with USP/China Pharmacopoeia Reference Standard Endotoxin. Lot No.: H0K354/150800-201601.

Reference:

United States Pharmacopoeia (USP), General Chapter <85>, Bacterial Endotoxins Test (BET) and China Pharmacopoeia <2020version>, volume 4, general rules 1143.

Operator	Sandy xiao	Date	2022.09.10
QA Administrator	Jane Lin	Date	2022.09.10

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.lalendotoxin.com>