



Endosafe<sup>®</sup>  
LAL Grade Glassware (T400)  
*Certificate of Quality*

Charles River Laboratories, Inc. 10 x 75 mm Borosilicate Glass Test Tubes without Caps, 50/Pack Lot Number 94911C has been depyrogenated in a dry heat oven using a validated process. The lot has been tested by Quality Control and meets the following test specifications:

Endotoxin Test Result: <0.005 EU/device

Interference Test Result: Pass

Expiration Date: Apr 2022

Checked By Paul August Date 03 Apr 2019

Reviewed By Donna Sief Date 03 Apr 2019

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA