



Endosafe[®] - PTS[™] Cartridges Certificate of Analysis



Reorder Code: PTS2001F Cartridge Lot #: 9726142 Expiration Date: Aug 2021



Calibration Code: 116347002627 RSE/CSE Ratio: 9 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 1-0.01 EU/mL Standard Curve Linearity: -0.998

Standard Curve Mean Reaction Times: 1.0 EU/mL 163 seconds

0.1 EU/mL 411 seconds

0.01 EU/mL 870 seconds

Archived Spike Concentration: 0.092 EU/mL Negative Control: Pass

This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst M. E. Gray

Date 19 Aug 2019

Reviewed By A. Smith

Date 19 Aug 2019

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA