

Endosafe<sup>®</sup>  
LAL Grade Glassware (T100)  
*Certificate of Quality*

Depyrogenated, 10 x 75 mm Tubes with Caps Lot Number P1605-17007 has been tested  
by Quality Control and meets the following test specification:

Endotoxin Test Result: < 0.005 EU/device

Expiration Date: Mar 2021

Product Code: T100

Checked By C Jesso Date 30 Nov 2017

Reviewed By KJenna Date 01 Dec 2017

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA