

Endosafe<sup>®</sup>  
LAL Grade Glassware (T100)  
*Certificate of Quality*

Depyrogenated, 10 x 75 mm Tubes with Caps Lot Number P1605-20008 has been tested  
by Quality Control and meets the following test specification:

Endotoxin Test Result: < 0.005 EU/device

Expiration Date: May 2024

Product Code: T100

Checked By *Christina Fortin* Date *21 Sep 2020*

Reviewed By *D. Somers* Date *21 Sep 2020*

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA