

BIOENDO



Xiamen Bioendo Technology Co., Ltd.

CERTIFICATE OF ANALYSIS

Catalog Number:	BT10	Lot Number:	23028109
Product:	β -glucans Blocker, 10ml/vial	Manufacture Date:	08-Feb-2023
		Expiration Date:	07-Feb-2025

TEST (Method)	SPECIFICATIONS	Results
Sterility	Negative	Negative
Endotoxin Level (EU/ml)	< 0.005	Pass
PPC Recovery(%)	50-150	104
Storage	2-30°C	

Operator *Sandy xiao*

Date *2023.02.12*

QA Administrator *Jane Lin*

Date *2023.02.12*

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.bioendo.com>