

BIOENDO



Xiamen Bioendo Technology Co., Ltd.

CERTIFICATE OF ANALYSIS

Catalog Number:	BY10	Lot Number:	22028108
Product:	10 mM Magnesium Buffer, 10ml/vial	Manufacture Date:	12-Feb-2022
		Expiration Date:	11-Feb-2024

Test (Method)	Specifications	Results
Sterility	Negative	Negative
Endotoxin Level (EU/ml)	< 0.005	Pass
PPC Recovery(%)	50-150	108

Operator *Sandy Xiao*

Date *2022.02.16*

QA Administrator *Jane Lin*

Date *2022.02.16*

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.lalendotoxin.com>