

Endosafe[®]
LAL Grade Glassware (T100)
Certificate of Quality

Depyrogenated, 10 x 75 mm Tubes with Caps Lot Number P16050216 has been tested
by Quality Control and meets the following test specification:

Endotoxin Test Result: < 0.015 EU/tube

Expiration Date: Mar 2020

Product Code: T100

Checked By *Jeetta Polite* Date *03 May 2016*

Reviewed By *A Chapman* Date *03 May 2016*

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA