

**Endosafe<sup>®</sup> - PTS<sup>™</sup> Cartridges**  
**Certificate of Analysis**



Reorder Code: PTS2005F Cartridge Lot #: 7677128 Expiration Date: Jun 2019



Calibration Code: 512329964424 RSE/CSE Ratio: 15 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 5-0.05 EU/mL Standard Curve Linearity: -0.991

Standard Curve Mean Reaction Times: 5.0 EU/mL 123 seconds  
0.5 EU/mL 358 seconds  
0.05 EU/mL 699 seconds

Archived Spike Concentration: 0.64 EU/mL Negative Control: Pass

**This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.**

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

**CAUTION: DO NOT FREEZE THE CARTRIDGES**

Qualified Analyst J. Hoj

Date 01 Aug 2017

Reviewed By M. Allen

Date 01 Aug 2017

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA