


BIOENDO



Xiamen Bioendo Technology Co., Ltd.

CERTIFICATE OF ANALYSIS

Catalog Number:	G520030	Lot Number:	22111163
Product:	Gel Clot Lyophilized Amebocyte Lysate Multi-test Vial (50 Tests/Vial)	Manufacture Date:	12-Nov-2022
Specifications	5.2 ml/vial	Expiration Date:	11-Nov-2024
Test Items	Description	Results	
Appearance	White lyophilized powder	Pass	
Sensitivity	Within 50% to 200% of specified sensitivity	Pass	
Buffer capacity	pH 6.0-8.0	Pass	
4h negative control	Negative	Pass	
Matched Endotoxin	Lot No. 22095149, 90EU/vial		
Storage	2-8°C		

The sensitivity of reagent is 0.03EU/ml when assay parallel with USP/China Pharmacopoeia Reference Standard Endotoxin. Lot No.: H0K354/150800-201601.

Reference:

United States Pharmacopoeia (USP), General Chapter <85>, Bacterial Endotoxins Test (BET) and China Pharmacopoeia <2020version>, volume 4, general rules 1143.

Operator	Sandy Xiao	Date	2022.11.17
QA Administrator	Jane Lin	Date	2022-11-17

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.lalendotoxin.com>