


# Luoyang Fudau Biotech Co., Ltd

## Certificate of Analysis

Inspection NO.: CJ- 202506001

<b>Product name</b>	Multi-wells cell culture plate	<b>Specification</b>	96wells, Non-TC treated
<b>Item No.</b>	C061096	<b>Production batch</b>	PE02250629
<b>Package</b>	Blister box, 1pcs/bag, 50bag/case	<b>Validity Period</b>	2028.05
<b>Material</b>	PS	<b>Report date</b>	2025.07.05
<b>Material Grade</b>	Meets the requirements of USP Class VI		
<b>Test items</b>	<b>Standard</b>		<b>Test Result</b>
<b>Exterior</b>	Glossy surface, no deformation, no black spots, cracks, flakes, miscellaneous materials, oil stains, fingerprints, bubbles, flashes, etc.		Meet the requirements
<b>Structure size</b>	Meet the requirements of the drawings		Meet the requirements
<b>Hydrophilia</b>	After TC treatment, the contact angle is not greater than 45°		NA
<b>Package</b>	Packing materials, quantity are correct, and packaging is intact		Meet the requirements
<b>Sterilize</b>	rradiation label should be red		Meet the requirements
<b>Bacterial endotoxin</b>	Not more than 0.1EU/ml		Meet the requirements
<b>In vitro cytotoxin</b>	Should meet the requirements of level 0		Meet the requirements
<b>Cell culture</b>	Cells are in good condition and morphology		NA
<b>DNase detection</b>	There should be no DNase		Meet the requirements
<b>RNase detection</b>	There should be no RNase		Meet the requirements
<b>Testing base</b>	 QW/01-2004		
<b>Conclusion</b>	Qualified		

Inspector

张梦圆

Reviewer

张丽

Approver

白静