



## CERTIFICATE OF ANALYSIS

<b>Catalog Number:</b>	G520030	<b>Lot Number:</b>	24071165
<b>Product:</b>	Gel Clot Lyophilized Amebocyte Lysate Multi-test Vial (50 Tests/Vial)	<b>Manufacture Date:</b>	24-Jul-2024
<b>Specifications</b>	5.2 ml/vial	<b>Expiration Date:</b>	23-Jul-2026
<b>Test Items</b>	<b>Description</b>	<b>Results</b>	
Appearance	White lyophilized powder	Pass	
Sensitivity	Within 50% to 200% of specified sensitivity	Pass	
Buffer capacity	pH 6.0-8.0	Pass	
4h negative control	Negative	Pass	
<b>Matched Endotoxin</b>	Lot No. 23045133, 90EU/vial; 24045124, 80EU/vial		
<b>Storage</b>	2-8°C		

The sensitivity of reagent is 0.03 EU/ml when assay parallel with USP/China Pharmacopoeia Reference Standard Endotoxin. Lot No.:H0K354/150800-201601.

### Reference:

United States Pharmacopoeia (USP), General Chapter <85>, Bacterial Endotoxins Test (BET) and China Pharmacopoeia <2020 version>, volume 4, general rules 1143.

Operator *Sandy xiao*

Date *2024.07.30*

QA Administrator *Jane Lin*

Date *2024.07.30*