



CERTIFICATE OF ANALYSIS

Catalog Number:	ECV2500V	Lot Number:	24015103
Product:	Endotoxin Challenge Vial	Manufacture Date:	14-Jan-2024
		Expiration Date:	13-Jan-2026

Bacterial strain *E.coli* O111: B4

Contents	Description	Catalog No.	Lot No.
Endotoxin Challenge Vial	≥ 2500 EU/vial	ECV2500V	24015103
TEST (Method)	SPECIFICATIONS	Results	
Endotoxin Potency (EU/vial)	≥ 2500 EU/vial	Pass	
Storage	2-8°C		

Reconstitute ECV with 1.0 ml Water for BET, mix vigorously for at least 15 minutes on a vortex mixer.

Operator *Sandy xiao*

Date *2024-01-21*

QA Administrator *Jane Lin*

Date *2024-01-21*