



Endosafe[®]
Endotoxin Indicators (EVV2.5M)
Certificate of Quality

Each vial of Endotoxin Indicator Lot #: EVV83161 Expiration Date: Nov 2020

contains approximately 250,000 ng of lipopolysaccharide (derived from *E.coli* Strain 055:B5).

The biological activity is approximately 2,500,000 Endotoxin Units (EU) of lipopolysaccharide, as determined by serial two-fold dilution at the endpoint with Gel Clot U.S. Reference Lysate Lot # 15.

The absence of assay interfering substances was demonstrated by acceptable spike recovery in the range of 50% to 200%.

Qualified Analyst *Tanna Aijf* Date 16 Nov 2018

Reviewed By *J Yukimi* Date 19 NOV 2018

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA