



Endosafe[®] - PTS[™] Cartridges Certificate of Analysis



Reorder Code: PTS20005F Cartridge Lot #: 6530171 Expiration Date: Jan 2022



Calibration Code: 418453304504 RSE/CSE Ratio: 9 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 0.5-0.005 EU/mL Standard Curve Linearity: -0.998

Standard Curve Mean Reaction Times: 0.5 EU/mL 184 seconds

0.05 EU/mL 448 seconds

0.005 EU/mL 933 seconds

Archived Spike Concentration: 0.065 EU/mL Negative Control: Pass

This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst Burkean Date 03 Feb 2020

Reviewed By Summit Date 03 Feb 2020

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA