


# BIOENDO



Xiamen Bioendo Technology Co., Ltd.

## CERTIFICATE OF ANALYSIS

<b>Catalog Number:</b>	ECV2500V	<b>Lot Number:</b>	24055139
<b>Product:</b>	Endotoxin Challenge Vial	<b>Manufacture Date:</b>	19-May-2024
		<b>Expiration Date:</b>	18-May-2026
<b>Bacterial strain</b>	<i>E.coli</i> O111: B4		
<b>Contents</b>	<b>Description</b>	<b>Catalog No.</b>	<b>Lot No.</b>
Endotoxin Challenge Vial	≥2500 EU/vial	ECV2500V	24055139
<b>TEST (Method)</b>	<b>SPECIFICATIONS</b>	<b>Results</b>	
Endotoxin Potency (EU/vial)	≥2500 EU/vial	Pass	
<b>Storage</b>	2-8°C		

Reconstitute ECV with 1.0 ml Water for BET, mix vigorously for at least 15 minutes on a vortex mixer.

Operator

Sandy xiao

Date

2024.05.25

QA Administrator

Jane Lin

Date

2024.05.25

Manufacturer Xiamen Bioendo Technology Co., Ltd.  
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China  
Website <http://www.bioendo.com>