



Endosafe<sup>®</sup>  
Endotoxin Indicators (EVV2K)  
Certificate of Quality

Each vial of Endotoxin Indicator Lot #: EVV81001 Expiration Date: Apr 2020

contains approximately 200 ng of lipopolysaccharide (derived from *E.coli* Strain 055:B5).

The biological activity is approximately 2,000 Endotoxin Units (EU) of lipopolysaccharide, as determined by serial two-fold dilution at the endpoint with Gel Clot U.S. Reference Lysate Lot # 15.

The absence of assay interfering substances was demonstrated by acceptable spike recovery in the range of 50% to 200%.

Qualified Analyst *Joanna Siegl* Date 18 Apr 2018

Reviewed By *KGenies* Date 18 Apr 2018

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA