



Endosafe[®] - PTS[™] Cartridges Certificate of Analysis



Reorder Code: PTS2001F Cartridge Lot #: 6595132 Expiration Date: Apr 2022



Calibration Code: 119955652529 RSE/CSE Ratio: 9 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 1-0.01 EU/mL Standard Curve Linearity: -0.996

Standard Curve Mean Reaction Times: 1.0 EU/mL 199 seconds

0.1 EU/mL 490 seconds

0.01 EU/mL 956 seconds

Archived Spike Concentration: 0.090 EU/mL Negative Control: Pass

This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst *[Signature]*

Date 06 Apr 2020

Reviewed By *[Signature]*

Date 07 Apr 2020

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA