


BIOENDO



Xiamen Bioendo Technology Co., Ltd.

CERTIFICATE OF ANALYSIS

Catalog Number: BT10 **Lot Number:** 22038114
Product: β -glucans Blocker, 10ml/vial **Manufacture Date:** 08-Mar-2022
Expiration Date: 07-Mar-2024

TEST (Method)	SPECIFICATIONS	Results
Sterility	Negative	Negative
Endotoxin Level (EU/ml)	< 0.005	Pass
PPC Recovery(%)	50-150	104
Storage	2-30°C	

Operator *Sandy xiao*

Date *2022-03-16*

QA Administrator *Jane Lin*

Date *2022-03-16*

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.lalendotoxin.com>