

**Endosafe<sup>®</sup> - PTS<sup>™</sup> Cartridges**  
**Certificate of Analysis**



Reorder Code: PTS20005F Cartridge Lot #: 7730166 Expiration Date: Aug 2019



Calibration Code: 424467153928 RSE/CSE Ratio: 10 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 0.5-0.005 EU/mL Standard Curve Linearity: -0.991

Standard Curve Mean Reaction Times: 0.5 EU/mL 244 seconds

0.05 EU/mL 607 seconds


0.005 EU/mL 1071 seconds

Archived Spike Concentration: 0.059 EU/mL Negative Control: Pass

**This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.**

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

**CAUTION: DO NOT FREEZE THE CARTRIDGES**

Qualified Analyst  Date 12 Apr 2018

Reviewed By  Date 12 Apr 2018

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA