



Endosafe[®]
LAL Grade Glassware (TL1000)
Certificate of Quality

Charles River Laboratories, Inc. 12 x 75 mm Borosilicate Glass Test Tubes without Caps, 50/Pack Lot Number 82741C has been depyrogenated in a dry heat oven using a validated process. The lot has been tested by Quality Control and meets the following test specifications:

Endotoxin Test Result: <0.005 EU/device

Interference Test Result: Pass

Expiration Date: Oct 2021

Checked By *Pal GA* Date 02 Oct 2018

Reviewed By *J. Yurkin* Date 04 Oct 2018

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA