



Endosafe<sup>®</sup>  
Endotoxin Indicators (EUV10K)  
*Certificate of Quality*

Each vial of Endotoxin Indicator Lot #: EVV94781 Expiration Date: Mar 2021

contains approximately 1,000 ng of lipopolysaccharide (derived from *E.coli* Strain 055:B5).

The biological activity is approximately 10,000 Endotoxin Units (EU) of lipopolysaccharide, as determined by serial two-fold dilution at the endpoint with Gel Clot U.S. Reference Lysate Lot # 15.

The absence of assay interfering substances was demonstrated by acceptable spike recovery in the range of 50% to 200%.

Qualified Analyst Christina Easton Date 22 Mar 2019

Reviewed By Domena Sijf Date 27 Mar 2019

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA