

PASS

CERTIFICATE OF ANALYSIS

Catalog Number:	G520030	Lot Number:	26051137
Product:	Gel Clot Lyophilized Amebocyte Lysate Multi-test Vial (50 Tests/Vial)	Manufacture Date:	16-May-2026
Specifications	5.2 ml/vial	Expiration Date:	15-May-2028
Test Items	Description	Results	
Appearance	White lyophilized powder	Pass	
Sensitivity	Within 50% to 200% of specified sensitivity	Pass	
Buffer capacity	pH 6.0-8.0	Pass	
4h negative control	Negative	Pass	
Matched Endotoxin	Lot No. 25075177, 80EU/vial;		
Storage	2-8°C		

The sensitivity of reagent is 0.03 EU/ml when assay parallel with USP/China Pharmacopoeia Reference Standard Endotoxin. Lot No.:R172R0/150800-201601.

Reference:

United States Pharmacopoeia (USP), General Chapter <85>, Bacterial Endotoxins Test (BET) and China Pharmacopoeia <2025 version>, volume 4, general rules 1143.

Operator Sandy Xiao

Date 2026.05.22

QA Administrator Jane Lin

Date 2026.05.22

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xijia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.bioendo.com>