


BIOENDO

Xiamen Bioendo Technology Co., Ltd.



CERTIFICATE OF ANALYSIS

Catalog Number:	BT10	Lot Number:	25018111
Product:	β -glucans Blocker, 10ml/vial	Manufacture Date:	16-Jan-2025
		Expiration Date:	15-Jan-2027

TEST (Method)	SPECIFICATIONS	Results
Sterility	Negative	Negative
Endotoxin Level (EU/ml)	< 0.005	Pass
PPC Recovery(%)	50-150	104
Storage	2-30°C	

Operator *Sandy xiao*

Date *2025.01.22*

QA Administrator *Jane Lin*

Date *2025.01.22*

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.bioendo.com>