



Endosafe[®]
Endotoxin Indicators (EVV1M)
Certificate of Quality

Each vial of Endotoxin Indicator Lot #: EVV84371 Expiration Date: Mar 2020

contains approximately 100,000 ng of lipopolysaccharide (derived from *E.coli* Strain 055:B5).

The biological activity is approximately 1,000,000 Endotoxin Units (EU) of lipopolysaccharide, as determined by serial two-fold dilution at the endpoint with Gel Clot U.S. Reference Lysate Lot # 15.

The absence of assay interfering substances was demonstrated by acceptable spike recovery in the range of 50% to 200%.

Qualified Analyst *Donna Singh* Date 02 Mar 2018

Reviewed By *Kayma* Date 05 Mar 2018

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA