



Endosafe[®]
LAL Grade Glassware (TL1200)
Certificate of Quality

Charles River Laboratories, Inc. 10 x 75 mm Borosilicate Glass Test Tubes with Screw Caps, 50/Pack Lot Number 84521C has been depyrogenated in a dry heat oven using a validated process. The lot has been tested by Quality Control and meets the following test specifications:

Endotoxin Test Result: <0.005 EU/device

Interference Test Result: Pass

Expiration Date: Feb 2021

Checked By C. Jones Date 27 Feb 2018

Reviewed By K. Jones Date 05 Mar 2018

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA