



Endosafe[®]
Positive Product Control (PC100)
Certificate of Quality

Endosafe[®] Positive Product (Inhibition)
Control Lot Number PCL3091 has been tested with
LAL Reagent Water and found to produce a positive
response (firm gel) at or within 60 minutes.

Expiration Date: Nov 2021

Qualified Analyst *[Signature]* Date 07 Nov 2019

Reviewed By *[Signature]* Date 07 NOV 2019

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA