



Endosafe[®] - PTS[™] Cartridges Certificate of Analysis



Reorder Code: PTS2005F Cartridge Lot #: 9731159 Expiration Date: Aug 2021



Calibration Code: 516338086802 RSE/CSE Ratio: 14 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 5-0.05 EU/mL Standard Curve Linearity: -0.999

Standard Curve Mean Reaction Times: 5.0 EU/mL 163 seconds

0.5 EU/mL 376 seconds

0.05 EU/mL 780 seconds

Archived Spike Concentration: 0.88 EU/mL Negative Control: Pass

This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst WEray

Date 20 Aug 2019

Reviewed By [Signature]

Date 20 Aug 2019

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA