



Endosafe[®]
0.1 M, 5.5 mL Tris Buffer (BT105)
Certificate of Quality

Endosafe 0.1 M, 5.5 mL Tris Buffer
Lot Number LC11001 has been tested by Quality
Control and found to contain: <0.005 EU/mL.

Expiration Date: Apr 2023

Qualified Analyst *Paul Aft* Date 15 Apr 2021

Reviewed By *J. J. J. J.* Date 15 Apr 2021

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA