


BIOENDO



Xiamen Bioendo Technology Co., Ltd.

CERTIFICATE OF ANALYSIS

Catalog Number:	ECV2500V	Lot Number:	23055145
Product:	Endotoxin Challenge Vial	Manufacture Date:	21-May-2023
		Expiration Date:	20-May-2025
Bacterial strain	<i>E.coli</i> O111: B4		

Contents	Description	Catalog No.	Lot No.
Endotoxin Challenge Vial	≥ 2500 EU/vial	ECV2500V	23055145
TEST (Method)	SPECIFICATIONS	Results	
Endotoxin Potency (EU/vial)	≥ 2500 EU/vial	Pass	
Storage	2-8°C		

Reconstitute ECV with 1.0 ml Water for BET, mix vigorously for at least 15 minutes on a vortex mixer.

Operator *Sandy xiao*

Date *2023.05.28*

QA Administrator *Jane Lin*

Date *2023.05.28*

Manufacturer Xiamen Bioendo Technology Co., Ltd.
Address 131 Xinjia Road, Haicang Xinyang Industrial Zone, Xiamen, China
Website <http://www.bioendo.com>