

Endosafe[®]
LAL Grade Glassware (T100)
Certificate of Quality

Depyrogenated, 10 x 75 mm Tubes with Caps Lot Number P1605-21007 has been tested
by Quality Control and meets the following test specification:

Endotoxin Test Result: < 0.005 EU/device

Expiration Date: May 2025

Product Code: T100

Checked By V. Morley Date 09 Sep 2021

Reviewed By g. yuerm Date 13 Sep 2021

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA