



Endosafe[®] - PTS[™] Cartridges Certificate of Analysis

Reorder Code: PTS20005F Cartridge Lot #: 1655148 Expiration Date: Jun 2023

Calibration Code: 424875844059 RSE/CSE Ratio: 10 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 0.5-0.005 EU/mL Standard Curve Linearity: -0.999

Standard Curve Mean Reaction Times: 0.5 EU/mL 248 seconds

0.05 EU/mL 572 seconds

0.005 EU/mL 1158 seconds

Archived Spike Concentration: 0.060 EU/mL Negative Control: Pass

This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst *Yvonne Pelito*

Date 16 Jun 2021

Reviewed By *C. Sin*

Date 22 Jun 2021

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA