



Endosafe<sup>®</sup>  
LAL Grade Glassware (T400)  
*Certificate of Quality*

Charles River Laboratories, Inc. 10 x 75 mm Borosilicate Glass Test Tubes without Caps, 50/Pack Lot Number 82111C has been depyrogenated in a dry heat oven using a validated process. The lot has been tested by Quality Control and meets the following test specifications:

Endotoxin Test Result: < 0.005 EU/device

Interference Test Result: Pass

Expiration Date: Jul 2021

Checked By *Donna Sijl* Date *01 Aug 2018*

Reviewed By *KGomes* Date *01 Aug 2018*

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA