



Endosafe[®]
Positive Product Control (PC100)
Certificate of Quality

Endosafe[®] Positive Product (Inhibition)
Control Lot Number PCM4701 has been tested with
LAL Reagent Water and found to produce a positive
response (firm gel) at or within 60 minutes.

Expiration Date: Mar 2022

Qualified Analyst Christina Foster Date 11 Mar 2020

Reviewed By Jyurm Date 12 mar 2020

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA