




## Endosafe<sup>®</sup> - PTS<sup>™</sup> Cartridges Certificate of Analysis

Reorder Code: PTS2001F Cartridge Lot #:  6726145 Expiration Date: Aug 2022

Calibration Code:  119053145593 RSE/CSE Ratio: 11 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 1-0.01 EU/mL Standard Curve Linearity: -0.999

Standard Curve Mean Reaction Times: 1.0 EU/mL 190 seconds  
0.1 EU/mL 444 seconds  
0.01 EU/mL 931 seconds

Archived Spike Concentration: 0.150 EU/mL Negative Control: Pass

**This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.**

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

**CAUTION: DO NOT FREEZE THE CARTRIDGES**

Qualified Analyst *Brunear* Date 17 Aug 2020

Reviewed By *J.E.* Date 18 Aug 2020

Charles River Laboratories, Inc.  
1023 Wappoo Road, Suite 43-B  
Charleston, SC 29407 USA