

**Endosafe[®] - PTS[™] Cartridges
Certificate of Analysis**

Reorder Code: PTS2005F Cartridge Lot #:  7798116 Expiration Date: Oct 2019

Calibration Code:  512641656191 RSE/CSE Ratio: 9 EU/ng RSE Lot #: H0K354

Archived Standard Curve Range: 5-0.05 EU/mL Standard Curve Linearity: -0.996


Standard Curve Mean Reaction Times: 5.0 EU/mL 126 seconds
0.5 EU/mL 371 seconds
0.05 EU/mL 816 seconds

Archived Spike Concentration: 0.81 EU/mL Negative Control: Pass

This lot of PTS Cartridges has been tested and meets Quality Control testing requirements for an archived curve, negative controls, and positive product control results.

Store cartridges at 2-25°C. Allow the unopened foil pouch to reach room temperature prior to opening. Cartridges should be used immediately once the foil pouch seal has been opened. Cartridges are for single-test use only.

CAUTION: DO NOT FREEZE THE CARTRIDGES

Qualified Analyst  Date 20 Dec 2017
Reviewed By M. Allen Date 20 Dec 2017

Charles River Laboratories, Inc.
1023 Wappoo Road, Suite 43-B
Charleston, SC 29407 USA